

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

JAN 12 2010

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130493</u>
Name <u>PATRICK SMITH</u>		ID# <u>6397</u> Date <u>31 DEC 09</u>
A	Agency <u>CLEAR AFS SECURITY FORCES</u>	Phone # <u>585-6093</u>
Instrument Location <u>BLDG 37, CLEAR AFS, AIL</u>		
Alco S/N <u>X124609</u>		Target Value <u>.081</u> High Pressure <u>1175</u>
B	Alco Test Values 1 <sup>st</sup> Alco <u>.076</u> 2 <sup>nd</sup> Alco <u>.077</u>	
Signature <u>[Signature]</u>		<u>COS</u> <u>1/13/10</u>
(OVER)		

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz

Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 20<sup>th</sup> day of January, 2010.

[Signature]  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

*of DataMaster cdm breath Test Instrument*

*State of Alaska*

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

DataMaster cdm S/N 130493 ✓

(CONTINUED FROM FRONT PAGE)

**C** TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130493 ✓

DECEMBER 31, 2009

CPB  
1/13/10

OPERATOR'S NAME:

SMITH

OPERATOR'S NUMBER: 6397

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: 1234567

DEPT/AGENCY: CLR4

CASE/REPORT: 09-1231

TEST TYPE: V

ALCO TARGET VALUE: .081 ✓

ALCO S/N: X124609 ✓

--- BREATH ANALYSIS ---

.081 ADJUSTED FOR 29.78 in ✓

ALCO TARGET .000 07:29

BLANK TEST .000 07:30

INTERNAL STANDARD VERIFIED 07:30

ALCO TV 29.78 in .076 ✓ 07:30

BLANK TEST .000 07:31

SUBJECT SAMPLE .000 07:32

BLANK TEST .000 07:32

ALCO TV 29.78 in .077 ✓ 07:33

BLANK TEST .000 07:33

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130493 ✓

DECEMBER 31, 2009 ✓

TIME 07:23

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 01/08/09

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 43c

BAROMETER: 29.78 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!\"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHIJ  
KLMNOPQRSTUVWXYZ[-\_`abdefghijklmno  
pqrstuvwxyz{|}~